



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 1/14/14-CD

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check. **REVIEWED** (35 days)
Complete this report whenever the instrument is serviced or repaired and whenever it is replaced or recalibrated.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.
By Carol Day at 12:02 pm, Feb 04, 2014

DATAMASTER SN 204162	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 01/08/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Gentry County Sheriff's Department		TIME OF INSPECTION 11:34 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 11:34
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50°C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

☒ INDICATOR LIGHTS

☒ SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc. LOT # 13001 EXP. DATE 03/07/2015

☒ SIMULATOR TEMP (34°C ± 0.2°C) 34.1 °C SIMULATOR SN G6788 EXP. DATE 10/16/2014

☒ CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 0.100%	TEST 2 101%	TEST 3 102%
---------------	-------------	-------------

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	3	(.10-.14)	0	(.15-.19)	1	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
(USE OTHER SIDE IF NECESSARY).

Operating within MODHSS Standards

INSPECTING OFFICER

SIGNATURE J	PRINT FULL NAME Justin S. Johnson
TYPE II PERMIT NUMBER/EXPIRATION DATE 220315 09/28/2014	TELEPHONE NUMBER (816) 387-2345

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).


The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013

The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204162
61.08/14

TESTING OFFICER:

JOHNSON, J/S

OFFICER I.D. # 513

PERMIT NUMBER: 200315

EXPIRATION DATE: 09-28-14

MISCELLANEOUS DATA:

---- SUPERVISOR MODE ----

BLANK TEST	.000	11:40
INTERNAL STANDARD	VERIFIED	11:40
EXTERNAL STANDARD	.100	11:40
BLANK TEST	.000	11:41
EXTERNAL STANDARD	.101	11:41
BLANK TEST	.000	11:42
EXTERNAL STANDARD	.102	11:43
BLANK TEST	.000	11:43

N = 3

SIM. = .1

AVG. = .101

Operator Signature

Printed on recycled paper with agri-based inks

CMSU 2208-02

BAC DataMaster

Evidence Ticket

01/05/19

11-541

DISTRIBUTION: 10-14th LA

COMPUTER: UCHY

上列的估計 (估計值) 是： 0.034

HEATERS
SAMPLE CHAMBER: 500

FLOID DE FECTORS ORBY

PUMP HIGH SPEED ONLY

DEFECTS: UC94

FILTERS: OKAY

QUARTZ STANDARD 0.00

CALL 1-800-4-A-ARMY

EXHIBIT 16.5

"#12% C+4,-./0123456789; :;<?@ABCD EFGHIJKLMNOPQRSTU VWXYZ[\]^_`ab cdefghijklmnopqr stuvwxyz{|}~"

Operator Signature

Printed on recycled paper with agri-based inks

CMSU 2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204162
01/08/14

ARREST TIME: 11:06
SUBJECT NAME:
JOE/JOHN
DOB: 01/01/80 SEX: M
STATE/D.L.: MO/123456
ARRESTING OFFICER:
JOHNSON/J/S
OFFICER I.D.: 513
TESTING OFFICER:
JOHNSON/J/S
OFFICER I.D.: 513
PERMIT NUMBER: 220815
EXPIRATION DATE: 09/28/14
MISCELLANEOUS DATA:
RT TEST

--- BREATH ANALYSTS ---

BLANK TEST	1.000	11:56
INTERNAL STANDARD	WEPIA (0.0)	11:56
RADIO INTERFERENCE		

Operator Signature

Printed on recycled paper with agn-based inks

CMSU 2208-02

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



JUSTIN S JOHNSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/28/2012

Number 220315

Expires 09/28/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)